

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100384

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 09/08/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.78: 0.080
LOT #: AG634001
EXPIRATION: 12/05/2018
TANK PRESSURE: 671 psi

BLANK TEST	0.000	23:54
INTERNAL STANDARD	VERIFIED	23:54
EXTERNAL STANDARD	0.075	23:54
BLANK TEST	0.000	23:55
EXTERNAL STANDARD	0.075	23:55
BLANK TEST	0.000	23:56
EXTERNAL STANDARD	0.075	23:56
BLANK TEST	0.000	23:57
EXTERNAL STANDARD	0.074	23:57

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber =	49.0°C	PASSED
Breath Tube =	46.1°C	PASSED

PUMP INFO

Flow Rate =	4.754 L/M	PASSED
-------------	-----------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------