

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100669
Start Date: 4/3/23
Date Received: 4/3/23

Analyst(s): C. OBryant

Reason for instrument evaluation

Touch screen not responding

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag | |

As Found Conditions:

Barometric Pressure (inHg): 29.58 Temperature (C): 21.9 Relative Humidity (%): 16

Linearity Test (attached)

0.020g/210L Mnf: <u>Airgas</u>	Lot: <u>AG217902</u>	Exp: <u>6/28/24</u>
0.080g/210L Mnf: <u>Airgas</u>	Lot: <u>AG 113402</u>	Exp: <u>5/14/23</u>
0.300g/210L Mnf: <u>Airgas</u>	Lot: <u>AG 217903</u>	Exp: <u>6/28/24</u>

Notes

- VOC (attached) Verify software version
 VOC Internal Standard value ≤ 3.00% Non-Drinking Subject Test (attached)

Condition of item: Good Needs Repair

- Ready for Adjustment
 Ready for Calibration

End Date: 4/4/23

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

LINEARITY CHECK


Alaska Department of Public Safety
DATAMASTER dmt: 100669

Date: 04/04/2023
Time: 11:08:25

SUPERVISOR NAME:
COLLEEN S OBRYANT
OPERATOR'S #: 6237

TARGET	AVERAGE	STD. DEV.
0.020	0.0192	0.0004
0.079	0.0760	0.0000
0.296	0.2916	0.0005

SLOPE = 0.988740
 R^2 = 0.999970



LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety
DATAMASTER dmt: 100669


Date: 04/04/2023
Time: 11:08:28

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
TARGET AT 29.57: 0.020
LOT #: AG217902
EXPIRATION: 06/28/2024
TANK PRESSURE: 1066 psi

BLANK TEST	0.000	11:09
INTERNAL STANDARD	VERIFIED	11:09
EXTERNAL STANDARD	0.019	11:09
BLANK TEST	0.000	11:10
EXTERNAL STANDARD	0.019	11:10
BLANK TEST	0.000	11:11
EXTERNAL STANDARD	0.019	11:12
BLANK TEST	0.000	11:12
EXTERNAL STANDARD	0.019	11:13
BLANK TEST	0.000	11:14
EXTERNAL STANDARD	0.020	11:14
BLANK TEST	0.000	11:15

Average = 0.0192
Std Dev = 0.0004



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
DATAMASTER dmt: 100669


Date: 04/04/2023
Time: 11:28:14

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.57: 0.079
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 268 psi

BLANK TEST	0.000	11:29
INTERNAL STANDARD	VERIFIED	11:29
EXTERNAL STANDARD	0.076	11:29
BLANK TEST	0.000	11:30
EXTERNAL STANDARD	0.076	11:30
BLANK TEST	0.000	11:31
EXTERNAL STANDARD	0.076	11:31
BLANK TEST	0.000	11:32
EXTERNAL STANDARD	0.076	11:32
BLANK TEST	0.000	11:33
EXTERNAL STANDARD	0.076	11:34
BLANK TEST	0.000	11:34

Average = 0.0760
Std Dev = 0.0000



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
DATAMASTER dmt: 100669


Date: 04/04/2023
Time: 11:36:36

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
TARGET AT 29.57: 0.296
LOT #: AG217903
EXPIRATION: 06/28/2024
TANK PRESSURE: 428 psi

BLANK TEST	0.000	11:37
INTERNAL STANDARD	VERIFIED	11:38
EXTERNAL STANDARD	0.291	11:38
BLANK TEST	0.000	11:39
EXTERNAL STANDARD	0.292	11:39
BLANK TEST	0.000	11:40
EXTERNAL STANDARD	0.292	11:40
BLANK TEST	0.000	11:41
EXTERNAL STANDARD	0.292	11:41
BLANK TEST	0.000	11:42
EXTERNAL STANDARD	0.291	11:42
BLANK TEST	0.000	11:43

Average = 0.2916
Std Dev = 0.0005



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100669

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/04/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.56: 0.079
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 250 psi

BLANK TEST	0.000	11:49
INTERNAL STANDARD	VERIFIED	11:49
Xq = 0.084	(0.52%)	
EXTERNAL STANDARD	0.077	11:49
X[1] = 0.0771 (-0.0015) (0.0030)		
BLANK TEST	0.000	11:50
EXTERNAL STANDARD	0.077	11:50
X[1] = 0.0769 (-0.0009) (0.0030)		
BLANK TEST	0.000	11:51
EXTERNAL STANDARD	0.077	11:51
X[1] = 0.0765 (-0.0014) (0.0023)		
BLANK TEST	0.000	11:52
EXTERNAL STANDARD	0.076	11:52
X[1] = 0.0763 (-0.0013) (0.0015)		
BLANK TEST	0.000	11:53
EXTERNAL STANDARD	0.076	11:53
X[1] = 0.0761 (-0.0018) (0.0014)		
BLANK TEST	0.000	11:54

Average = 0.0766
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8 °C (44.0 °C - 52.0 °C)	PASSED
Breath Tube = 46.0 °C (38.0 °C - 50.0 °C)	PASSED

SETTINGS

Lamp Voltage = 1.64 V (0.98 V - 2.30 V)	PASSED
Cooler Voltage = 1.61 V (0.98 V - 2.28 V)	PASSED
Bias Voltage = 80 V (48 V - 112 V)	PASSED
Chopper Freq = 523 Hz (475 Hz - 575 Hz)	PASSED
Barometer = 29.6 in	

PUMP INFO

Flow Rate = 4.401 L/M (3.500 L/M - 6.500 L/M)	PASSED
--	--------

DETECTOR INFO


PUMP ON (0.002 V <= 0.010 V)	PASSED
PUMP OFF (0.001 V <= 0.010 V)	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.084	(0.10%)	(0.00% - 4.00%)	PASSED
------------	---------	-----------------	--------



36.00

32.00

28.00

24.00

20.00

16.00

12.00

8.00

4.00

0.00

30

25

20

10

5

0

