

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100689

Received From Anchorage PD

Date 02/11/21

Analyst Derch Walton

1. Reason for instrument return:

Previously monitored for various status messages, standard out of range, and static analysis error

Not returned with case DM 02/11/21

2. Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database
3. Ensure change in instrument status form is in JusticeTrax
4. Upload would not turn on at check in. Replaced on/off switch boots up now.
5. Update agency information and polling status in DM Host Calling Program DM 02/11/21
6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)