## Alaska Scientific Crime Detection Laboratory

# DataMaster Check In Form Issued: 1/13/2017 Version: DMCIF 2017 R0 Effective: 1/17/2017 Status: Active Instrument S/N /00674 Received From Intoximeters Analyst Derch Walton 1. Reason for instrument return: Returbished by factory for screen Freezing 2. Change instrument and case location in JusticeTrax Breath Alcohol Database 3. Ensure change in instrument status form is in JusticeTrax Instrument not in field IM 10/23/20 4. Upload 5. Update agency information and polling status in DM Host Calling Program 6. Check for VOCs $\square$

#### From Intoximeters

7. Intoximeters Repair Paperwork (attached)

### Intoximeters

Experience · Service · Integrity

### Service Record

SRO Number: SRO-042070 Received Date: 09/16/2020 Completion Date: 10/08/2020

SRO Type: REPAIR SRO Description: DMT Repair/Preventive Maint

Customer ID: C000003651 Customer PO: 09162020-Walton

Contact: Derek Walton Email: derek.walton@alaska.gov

Ship To: Units on SRO

Alaska Department of Public Safety Procurement & Supply Section 4805 Dr. MLK Jr. Avenue Anchorage AK 99507 LEGDMT LEGACY DMT

Service Line:	**All Instruments are Calibrated to Factory Specifications**
Unit Type:	(LEGDMT) LEGACY DMT
Serial Number:	50100674
Reason:	DMT Mechanical or Physical
	Display
Reason Notes:	Customer Reported Issue: Screen not working – becomes unresponsive after some use
Resolution:	Resolder
	Resolder
	Reflowed solder on the piggyback board. Reseated connectors

Parts:	<u>Oty</u>	Part Number	Part Description				
	ı	001199-00	DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN				
Reason:	Preventive Maintenance planned						
Reason Notes:	Preventive Maintenance (90-8000-00)						
	Drygas Regulator Assy Replacement @ \$225.00						
Resolution:	Prev M						
PrevMain							
Resolution Notes:	Recalibrated touch Adjusted tech ser Adjusted RFI sen Cleaned & adjust Replaced filter was Replaced filter was Replaced choppe Cleaned, lubricat Installed TB-30 I Installed torx ser Cleaned and adjust Instrument Calibration Leak checked - 2  Parts:  001199-00 3.3  59-0008-00 Co  Cal. solution (0.2  Calibration solutions and control of the control of	ch screen een settings sitivity ed filter wheel solenoid heel solenoid mounting screws motor adjustment lock down screw heel motor pinion r wheel collar ed, & adjusted filter wheel movem DMT Digital Pots Mod ews and seal tape to sample chamb sted the 5-way valve rated to Factory Specifications 4 hours  volt regulator llar-chopper wheel  00) Lot#20160 Bottle#340 (during on (0.200) Lot#20160 Bottle#114 103, Tank 050, Exp. 02-12-2022	ent parts er g repair) 7				
	2	59-0008-00	DMT HRDW COLLAR-CHOPPER WHEEL 31433				
	1	90-8000-00	CONTRACT PREV MAINT DMT				

## Order Packing Slip Report

Intoximeters
Experience · Service · Integrity

Pack Date: 10/15/2020

10/15/2020 9:29:47 AM

Packing Slip: 213338

SRO: SRO-042070

From: Intoximeters, Inc

2081 Craig Road St Louis MO 63146

**USA** 

Customer: C000003651

Alaska Department of Public Safety

4805 Dr MLK Jr Ave Anchorage AK 99507

USA

Ship To: 2

Alaska Department of Public Safety Procurement & Supply Section 4805 Dr. MLK Jr. Avenue Anchorage AK 99507

USA

Customer PO: 09162020-Walton

Ship Via: FDX

FED EX

Weight: 38.00

Delivery Terms: 2D

FOB: ORIGIN

# Packages: 1

**Customer Ship Account:** 

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity			
1	LEGDMT	LEGACY	DMT		50100674			
1	001199-00	EA		1.00	0.00			
DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN								
2	59-0008-00	EA		2.00	0.00			
DMT HRDW COLLAR-CHOPPER WHEEL 31433								
3	90-8000-00	EA		1.00	0.00			
CONTRACT PREV MAINT DMT								

#### CALIBRATION FACTORS

\_\_\_\_\_\_

Alaska Department of Public Safety

DATAMASTER dmt: 100674

Date: 10/05/2020 Time: 07:20:02

OPERATOR NAME: INTOX I METERS CERTIFICATE #: TEST

Ca = 0.2000

Cu		0.2000					
CAL	=	1.000188	0.800	<=	CAL	<	1.200
b1	=	0.0006	0.0000	<=	b1	<	0.0040
b2	-	0.0047	0.0010	<=	b2	<	0.0100
b3	=	0.0005	0.0000	<=	b3	<	0.0040
Χq	==	0.0973	0.0500	<=	Χq	<	0.2000
a21	=	1.119332	1.050	<=	a21	<	1.300
a31	=	0.417048	0.300	<=	a31	<	0.800