

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100670

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/15/2025

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.10: 0.078

LOT #: AG528101

EXPIRATION: 10/08/2027

TANK PRESSURE: 1207 psi

|                   |              |       |
|-------------------|--------------|-------|
| BLANK TEST        | 0.000        | 09:26 |
| INTERNAL STANDARD | VERIFIED     | 09:26 |
| EXTERNAL STANDARD | 0.078        | 09:26 |
| BLANK TEST        | 0.000        | 09:27 |
| EXTERNAL STANDARD | 0.080        | 09:27 |
| BLANK TEST        | 0.000        | 09:28 |
| EXTERNAL STANDARD | 0.079        | 09:28 |
| BLANK TEST        | 0.000        | 09:29 |
| EXTERNAL STANDARD | INTERFERENCE | 09:29 |

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED

Breath Tube = 48.1°C PASSED

### PUMP INFO

Flow Rate = 4.624 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

### FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only JW 12/22/25