

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 7/6/2025

Instrument #: 100364

Instrument Location: Nome Police Department

Supervisor Name and Agency: Officer Thomas Broady / Nome Police Department

Describe the Status Change or Issue with the Instrument:

Rebooted instrument per discussion with Derek Walton about instrument communication issues with the crime lab. Occurred 07/06/2025 at approximately 0203 hours.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

-----  
**For Use by SCDL**

**Additional Notes**

Reboot was to attempt to resolve lack of instrument communication with the lab. The reboot did not resolve this issue. DJW 07/08/2025

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor