



State of Alaska  
Department of Public Safety  
**Scientific Crime Detection Laboratory**

4805 Dr. Martin Luther King Jr. Ave.  
Anchorage, Alaska 99507  
David Kanaris, Chief, Forensic Laboratories

Breath Alcohol Calibration Report

Report Date: 10/14/2022

Calibration Item: Datamaster DMT

T690BCQFWXPFY1

Serial Number: 100683

Date Received: Not Recorded

Condition of item: Not Recorded

Method Used: DataMaster DMT Calibration/Current BAPM

Calibration Range: 0.020 g/210L - 0.300 g/210L

**No As-Found Data Collected**

**Start Date 10/12/2022 - As Left - End Date 10/13/2022**

As Left Conditions - Barometric Pressure (inHg): 29.54 Temperature (C): 21.7 Relative Humidity (%): 19

As-Left Data									
Nominal Values (g/210L)	Target Value at 29.54 in Hg	Lot Number	Results (g/210L)					Mean (g/210L)	Uncertainty (g/210L)
0.020	0.020	AG100401	0.020	0.020	0.020	0.020	0.020	0.020	+/- 0.002
0.040	0.039	AG111710	0.039	0.039	0.039	0.039	0.040	0.039	+/- 0.003
0.080	0.079	AG135004	0.079	0.079	0.078	0.078	0.078	0.078	+/- 0.003
0.100	0.099	AG108404	0.098	0.098	0.098	0.099	0.098	0.098	+/- 0.003
0.150	0.148	AG134205	0.147	0.146	0.147	0.147	0.147	0.147	+/- 0.004
0.300	0.296	AG130502	0.291	0.291	0.291	0.291	0.291	0.291	+/- 0.008

No As Found Evaluation performed due to last Evaluation completed prior to Breath Alcohol Procedure Manual version 5.0.

Calibration is performed using dry gas standards traceable to NIST. Standards are sourced from an external provider accredited to ISO/IEC 17025 and ISO 17034 standards. Internal instrument barometer readings are traceable to NIST. External barometers are calibrated by an external calibration provider accredited to the ISO/IEC 17025 standard.

The As-Left results listed above include the estimated measurement uncertainty with a coverage factor k=2 and an expanded coverage probability of 95%. The estimated uncertainty applies to the calibration process and not to subject breath samples.

**This report applies only to the item being calibrated and shall not be reproduced, except in full, without written approval from the laboratory.**

A handwritten signature in black ink that reads "Derek J. Walton".

Analyst: Derek Walton  
[derek.walton@alaska.gov](mailto:derek.walton@alaska.gov)

Technically Reviewed by: Carlie Glaister on 10/14/2022

# Alaska Scientific Crime Detection Laboratory

## DataMaster Calibration Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100683 Analyst: Derek Watto Start Date: 10/12/22

- Diagnostic Test (attached)  Internal Standard value ≤ 3.00%
- Adjust voltages, if needed
- Lamp (2.60 V Maximum)  Chopper Frequency
- Bias  Detector Voltage
- Cooler (2.00 V Maximum) (-0.100V to 0.100V)
- Set RFI Sensitivity
- Adjust internal barometer reading to match NIST-traceable barometer
- Adjust volume setting, if needed (1.50 L +/- 0.10 L)
- Print Technician Screen (attached)
- Print Calibration Factors (attached)
- Minimum Volume Test 1.5 L (1.3L to 1.7L)

### As Left Conditions:

Barometric Pressure (in Hg): 29.54 Temperature (C): 21.7 Relative Humidity (%) 19

### Linearity Test (attached)

0.020g/210L Mnf:	Airgas	Lot:	A6100401	Exp:	01/04/23
0.040g/210L Mnf:	Airgas	Lot:	A6111710	Exp:	04/27/23
0.080g/210L Mnf:	Airgas	Lot:	A6135004	Exp:	12/16/23
0.100g/210L Mnf:	Airgas	Lot:	A6108404	Exp:	03/25/23
0.150g/210L Mnf:	Airgas	Lot:	A6134205	Exp:	12/08/23
0.300g/210L Mnf:	Airgas	Lot:	A6130502	Exp:	11/01/23

### Functional Tests (attached)

- Incomplete  Invalid
- Maximum Attempts Exceeded  Radio Frequency Interference (RFI)
- Interference (Volatiles)

Acetone Mnf: Lifeloc Lot: 21DP41158 Level: 0.100 Exp: 04/29/23

Simulator S/N: SD1200 Verification Performed?

Isopropanol Mnf: Lifeloc Lot: 21KP41172 Level: 0.100 Exp: 11/09/23

Simulator S/N: DRU176 Verification Performed?

Neg. Control Mnf: Lifeloc Lot: 21KP41172 Level: 0.100 Exp: 11/09/23

Simulator S/N: DR41b2 Verification Performed?

### Calibrated External Thermometer:

Thermometer S/N: 210958384 Calibration Due Date: 11/09/23

- Verify stylus is calibrated  Non-Drinking Subject Test
- Reset Options  Upload
- Clear Memory  Remove Tank Prior to Shutdown
- Verification of Calibration (VOC) Calibration End Date: 10/13/22

# DIAGNOSTIC RECORD

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/12/2022  
Time: 12:24:36

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

## TEMPERATURES

Sample Chamber = 48.8 °C PASSED  
(44.0 °C - 52.0 °C)  
Breath Tube = 48.1 °C PASSED  
(38.0 °C - 50.0 °C)

## SETTINGS

Lamp Voltage = 1.95 V PASSED  
(1.17 V - 2.73 V)  
Cooler Voltage = 1.89 V PASSED  
(1.13 V - 2.65 V)  
Bias Voltage = 79 V PASSED  
(47 V - 111 V)  
Chopper Freq = 544 Hz PASSED  
(475 Hz - 575 Hz)  
Barometer = 29.6 in

## PUMP INFO

Flow Rate = 4.128 L/M PASSED  
(3.500 L/M - 6.500 L/M)

## DETECTOR INFO

PUMP ON PASSED  
(0.001 V <= 0.010 V)  
PUMP OFF PASSED  
(0.002 V <= 0.010 V)

## FILTER INFO

Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

INTERNAL STANDARD ✓ PASSED  
 $X_q = 0.104 \quad 1.09\%$  (0.00% - 4.00%)

Technician Screen

-----  
DATAMASTER dmt: 100683

Date: 10/12/2022  
Time: 12:56 PM

Temperatures (°C)  
Sample Cell: 48.9 °C  
Breath Tube: 46.6 °C

Current Barometer 29.6 in  
Volume (Ltr) 1.43

Settings  
Lamp: 1.76 V  
Bias: 79 V  
Cooler: 1.79 V  
Chopper: 527 Hz

Voltages (V)  
Flow: 1.02 V  
Detector: -0.055 V

## CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 09/01/2022  
Time: 11:28:48

OPERATOR NAME:  
TATIANA MALNOSKE  
CERTIFICATE #: 10437

Ca = 0.1000  
CAL = 0.965272 0.800 <= CAL < 1.200  
b1 = 0.0011 0.0000 <= b1 < 0.0040  
b2 = 0.0076 0.0010 <= b2 < 0.0100  
b3 = 0.0005 0.0000 <= b3 < 0.0040  
Xq = 0.1052 0.0500 <= Xq < 0.2000  
a21 = 1.200662 1.050 <= a21 < 1.300  
a31 = 0.521225 0.300 <= a31 < 0.800

40.00

36.00

32.00

28.00

24.00

20.00

16.00

12.00

8.00

4.00

0.00



— Alcohol (G/210L)  
 - - - Flow Rate (L/M)  
 - - - Min. Flow Rate (L/M)

## SUBJECT TEST

Alaska Department of Public Safety  
 DATAMASTER dmt: 100683

Date: 10/12/2022  
 Time: 12:57:49

SUBJECT NAME:  
 TEST TEST  
 LICENSE #: XXXX

OPERATOR NAME:  
 DEREK WALTON  
 OPERATOR'S #: 9508

DEPT/AGENCY:  
 MINIMUM VOLUME  
 CASE/REPORT: XXXX  
 TEST TYPE: TEST

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.54: 0.079

LOT #: AG135004

EXPIRATION: 12/16/2023

DIAGNOSTIC CHECK	PASSED	13:03
BLANK TEST	0.000	13:03
INTERNAL STANDARD	VERIFIED	13:03
EXTERNAL STANDARD	0.079	13:03
BLANK TEST	0.000	13:04
SUBJECT SAMPLE (Vol=1.51L)	0.000	13:05
BLANK TEST	0.000	13:06
EXTERNAL STANDARD	0.079	13:06
BLANK TEST	0.000	13:07

### LINEARITY CHECK

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:14:06

SUPERVISOR NAME:  
DEREK WALTON  
OPERATOR'S #: 9508

TARGET AVERAGE STD. DEV.

0.020	0.0200	0.0000
0.039	0.0392	0.0004
0.079	0.0784	0.0005
0.099	0.0982	0.0004
0.148	0.1468	0.0004
0.296	0.2910	0.0000

SLOPE = 0.981341  
 $R^2$  = 0.999982



DMW

# LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:14:08

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020  
TARGET AT 29.53: 0.020  
LOT #: AG100401  
EXPIRATION: 01/04/2023  
TANK PRESSURE: 545 psi

BLANK TEST	0.000	13:15
INTERNAL STANDARD	VERIFIED	13:15
EXTERNAL STANDARD	0.020	13:15
BLANK TEST	0.000	13:16
EXTERNAL STANDARD	0.020	13:17
BLANK TEST	0.000	13:17
EXTERNAL STANDARD	0.020	13:18
BLANK TEST	0.000	13:19
EXTERNAL STANDARD	0.020	13:19
BLANK TEST	0.000	13:20
EXTERNAL STANDARD	0.020	13:20
BLANK TEST	0.000	13:21

Average = 0.0200  
Std Dev = 0.0000



# LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:23:14

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.040  
TARGET AT 29.52: 0.039  
LOT #: AG111710  
EXPIRATION: 04/27/2023  
TANK PRESSURE: 127 psi

BLANK TEST	0.000	13:24
INTERNAL STANDARD	VERIFIED	13:24
EXTERNAL STANDARD	0.039	13:24
BLANK TEST	0.000	13:25
EXTERNAL STANDARD	0.039	13:25
BLANK TEST	0.000	13:26
EXTERNAL STANDARD	0.039	13:26
BLANK TEST	0.000	13:27
EXTERNAL STANDARD	0.039	13:28
BLANK TEST	0.000	13:28
EXTERNAL STANDARD	0.040	13:29
BLANK TEST	0.000	13:30

Average = 0.0392  
Std Dev = 0.0004



# LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:33:08

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.52: 0.079  
LOT #: AG135004  
EXPIRATION: 12/16/2023  
TANK PRESSURE: 1096 psi

BLANK TEST	0.000	13:34
INTERNAL STANDARD	VERIFIED	13:34
EXTERNAL STANDARD	0.079	13:34
BLANK TEST	0.000	13:35
EXTERNAL STANDARD	0.079	13:35
BLANK TEST	0.000	13:36
EXTERNAL STANDARD	0.078	13:37
BLANK TEST	0.000	13:38
EXTERNAL STANDARD	0.078	13:38
BLANK TEST	0.000	13:39
EXTERNAL STANDARD	0.078	13:39
BLANK TEST	0.000	13:40

Average = 0.0784  
Std Dev = 0.0005



# LINEARITY CHECK: Control Test 4

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:41:35

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.100  
TARGET AT 29.52: 0.099  
LOT #: AG108404  
EXPIRATION: 03/25/2023  
TANK PRESSURE: 191 psi

BLANK TEST	0.000	13:42
INTERNAL STANDARD	VERIFIED	13:42
EXTERNAL STANDARD	0.098	13:43
BLANK TEST	0.000	13:44
EXTERNAL STANDARD	0.098	13:44
BLANK TEST	0.000	13:45
EXTERNAL STANDARD	0.098	13:45
BLANK TEST	0.000	13:46
EXTERNAL STANDARD	0.099	13:46
BLANK TEST	0.000	13:47
EXTERNAL STANDARD	0.098	13:47
BLANK TEST	0.000	13:48

Average = 0.0982  
Std Dev = 0.0004



# LINEARITY CHECK: Control Test 5

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:48:48

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.150  
TARGET AT 29.51: 0.148  
LOT #: AG134205  
EXPIRATION: 12/08/2023  
TANK PRESSURE: 1099 psi

BLANK TEST	0.000	13:49
INTERNAL STANDARD	VERIFIED	13:50
EXTERNAL STANDARD	0.147	13:50
BLANK TEST	0.000	13:51
EXTERNAL STANDARD	0.146	13:51
BLANK TEST	0.000	13:52
EXTERNAL STANDARD	0.147	13:52
BLANK TEST	0.000	13:53
EXTERNAL STANDARD	0.147	13:53
BLANK TEST	0.000	13:54
EXTERNAL STANDARD	0.147	13:54
BLANK TEST	0.000	13:55

Average = 0.1468  
Std Dev = 0.0004



# LINEARITY CHECK: Control Test 6

Alaska Department of Public Safety  
DATAMASTER dmt: 100683


Date: 10/12/2022  
Time: 13:56:02

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300  
TARGET AT 29.50: 0.296  
LOT #: AG130502  
EXPIRATION: 11/01/2023  
TANK PRESSURE: 234 psi

BLANK TEST	0.000	13:57
INTERNAL STANDARD	VERIFIED	13:57
EXTERNAL STANDARD	0.291	13:57
BLANK TEST	0.000	13:58
EXTERNAL STANDARD	0.291	13:58
BLANK TEST	0.000	13:59
EXTERNAL STANDARD	0.291	13:59
BLANK TEST	0.000	14:00
EXTERNAL STANDARD	0.291	14:01
BLANK TEST	0.000	14:02
EXTERNAL STANDARD	0.291	14:02
BLANK TEST	0.000	14:03

Average = 0.2910  
Std Dev = 0.0000



40.00

36.00

32.00

28.00

24.00

20.00

16.00

12.00

8.00

4.00

0.00



SUBJECT TEST												
Alaska Department of Public Safety												
DATAMASTER dmt: 100683												
Date:	10/12/2022											
Time:	14:07:56											
SUBJECT NAME:												
TEST TEST												
LICENSE #:	XXXX											
OPERATOR NAME:	DEREK WALTON											
OPERATOR'S #:	9508											
DEPT/AGENCY:												
INCOMPLETE												
CASE/REPORT:	XXXX											
TEST TYPE:	TEST											
EXTERNAL STANDARD INFORMATION												
NOMINAL:	0.080											
TARGET AT:	29.48:											
0.079												
LOT #:	AG135004											
EXPIRATION:	12/16/2023											
DIAGNOSTIC CHECK	PASSED											
BLANK TEST	0.000											
INTERNAL STANDARD	VERIFIED											
EXTERNAL STANDARD	0.079											
BLANK TEST	0.000											
SUBJECT SAMPLE	INCOMPLETE											
	14:09											
	14:09											
	14:09											
	14:09											
	14:10											
	14:12											
	16.00											
	12.00											
	8.00											
	4.00											
	0.00											
JM												
0	10	20	30	40	50	60	70	80	90	100	110	120

40.00

36.67

33.33

30.00

26.67



SUBJECT TEST	
-- Alaska Department of Public Safety	
DATAMASTER dmt: 100683	
Date: 10/13/2022	
Time: 11:17:35	
SUBJECT NAME:	
TEST TEST	
LICENSE #: XXXX	
OPERATOR NAME:	
DEREK WALTON	
OPERATOR'S #: 9508	
DEPT/AGENCY:	
INVALID	
CASE/REPORT: XXXX	
TEST TYPE: TEST	
EXTERNAL STANDARD INFORMATION	
NOMINAL: 0.080	
TARGET AT 29.55: 0.079	
LOT #: AG135004	
EXPIRATION: 12/16/2023	
DIAGNOSTIC CHECK	PASSED 11:19
BLANK TEST	0.000 11:19
INTERNAL STANDARD	VERIFIED 11:19
EXTERNAL STANDARD	0.079 11:19
BLANK TEST	0.000 11:20
SUBJECT SAMPLE	INVALID 11:21

0.00

3.33

6.67

10.00

13.33

16.67

25

20

15

10

5


0

40.00

36.00

32.00

0.00

**SUBJECT TEST**

Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/12/2022  
Time: 14:19:47

SUBJECT NAME:  
TEST TEST  
LICENSE #: XXXX

OPERATOR NAME:

DEREK WALTON

OPERATOR'S #: 9508

DEPT/AGENCY:  
MAXIMUM ATTEMPTS

CASE/REPORT: XXXX  
TEST TYPE: TEST

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.080

TARGET AT 29.48: 0.079

LOT #: AG135004

EXPIRATION: 12/16/2023

DIAGNOSTIC CHECK	PASSED	14:20
BLANK TEST	0.000	14:20
INTERNAL STANDARD	VERIFIED	14:21
EXTERNAL STANDARD	0.079	14:21
BLANK TEST	0.000	14:22
SUBJECT SAMPLE	XXX	14:23

Maximum Attempts Exceeded

SUBJECT TEST			
Alcohol (g/210L)			
Flow Rate (L/M)			
Min. Flow Rate (L/M)			
	36.00		
		32.00	
SUBJECT NAME:			
TEST TEST			
LICENSE #: XXXX			
OPERATOR NAME:			
DEREK WALTON			
OPERATOR'S #: 9508			
DEPT/AGENCY:			
RADIO FREQUENCY			
CASE/REPORT: XXXX			
TEST TYPE: TEST			
EXTERNAL STANDARD INFORMATION			
NOMINAL: 0.080			
TARGET AT 29.57: 0.079			
LOT #: AG135004			
EXPIRATION: 12/16/2023			
DIAGNOSTIC CHECK	PASSED	11:48	
BLANK TEST	0.000	11:48	
INTERNAL STANDARD	VERIFIED	11:48	
EXTERNAL STANDARD	0.078	11:48	
BLANK TEST	0.000	11:49	
RFI DETECTED			
		16.00	
			12.00
			8.00
			4.00
			0.00
			30.00
			25.00
			20.00
			15.00
			10.00
			5.00
			0.00

## SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/13/2022  
Time: 11:46:33

SUBJECT NAME:  
TEST TEST  
LICENSE #: XXXX

OPERATOR NAME:  
DEREK WALTON  
OPERATOR'S #: 9508

DEPT/AGENCY:  
RADIO FREQUENCY  
CASE/REPORT: XXXX  
TEST TYPE: TEST

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.57: 0.079  
LOT #: AG135004  
EXPIRATION: 12/16/2023

DIAGNOSTIC CHECK	PASSED	11:48
BLANK TEST	0.000	11:48
INTERNAL STANDARD	VERIFIED	11:48
EXTERNAL STANDARD	0.078	11:48
BLANK TEST	0.000	11:49

## RFI DETECTED

40.00

36.00

32.00

28.00

24.00

20.00


16.00

12.00

8.00

4.00

0.00

**SUBJECT TEST**

Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/13/2022  
Time: 12:02:38

**SUBJECT NAME:**

TEST TEST  
LICENSE #: XXXX

**OPERATOR NAME:**

DEREK WALTON  
OPERATOR'S #: 9508

**DEPT/AGENCY:**

INTERFERENCE ACETONE  
CASE/REPORT: XXXX  
TEST TYPE: TEST

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.080

TARGET AT 29.57: 0.079

LOT #: AG135004

EXPIRATION: 12/16/2023

**DIAGNOSTIC CHECK**

BLANK TEST

PASSED

12:03

0.000

VERIFIED

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

12:03

0.080

0.079

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

0.078

Interference Detected

*JM*

40.00

36.00

32.00

28.00

24.00

20.00



16.00

12.00


8.00

4.00

0.00



40.00



100683 2022.10.13 Calibration Page 20 of 21

**SUBJECT TEST**

Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/13/2022  
Time: 12:19:35

**SUBJECT NAME:**

TEST TEST  
LICENSE #: XXXX

**OPERATOR NAME:**

DEREK WALTON  
OPERATOR'S #: 9508

**DEPT/AGENCY:**

INTERFERENCE NEGATIVE  
CASE/REPORT: XXXX  
TEST TYPE: TEST

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.080  
TARGET AT 29.58: 0.079  
LOT #: AG135004  
EXPIRATION: 12/16/2023

DIAGNOSTIC CHECK	PASSED	12:20
BLANK TEST	0.000	12:20
INTERNAL STANDARD	VERIFIED	12:20
XQ = 0.105 (0.25%)		
EXTERNAL STANDARD	0.078	12:21
X[1] = 0.0783 (-0.0024)	(0.0027)	
BLANK TEST	0.000	12:21
SUBJECT SAMPLE (V <sub>o1</sub> =1.86L)	0.094	12:22
X[1] = 0.0941 (-0.0024)	(0.0007)	
BLANK TEST	0.000	12:23
EXTERNAL STANDARD	0.078	12:23
X[1] = 0.0782 (-0.0021)	(0.0022)	
BLANK TEST	0.000	12:24

filter 2 average: -0.0024  
filter 3 average: 0.00628

DM

40.00

36.00

32.00

28.00

24.00

20.00

16.00

12.00

8.00

4.00

0.00

25


20

15

10

5

0



## SUBJECT TEST

Alaska Department of Public Safety  
DATAMASTER dmt: 100683

Date: 10/13/2022  
Time: 15:40:23

SUBJECT NAME:

TEST TEST  
LICENSE #: XXXX  
OPERATOR NAME:  
DEREK WALTON  
OPERATOR'S #: 9508

DEPT/AGENCY:

NON-DRINKING SUBJECT  
CASE/REPORT: XXXX  
TEST TYPE: TEST

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.65: 0.079

LOT #: AG135004

EXPIRATION: 12/16/2023

DIAGNOSTIC CHECK	PASSED	15:41
BLANK TEST	0.000	15:41
INTERNAL STANDARD	VERIFIED	15:41
EXTERNAL STANDARD	0.079	15:41
BLANK TEST	0.000	15:42
SUBJECT SAMPLE (Vol=3.17L)	0.000	15:43
BLANK TEST	0.000	15:44
EXTERNAL STANDARD	0.079	15:44
BLANK TEST	0.000	15:45