

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7 May 2015 Instrument Serial Number: 100421

Instrument Location: JBER, AK Bldg 8517

Supervisor Name and Agency: Scott Gardner 6222 JBER4

Describe the Status Change or Issue with the Instrument:

(6) six status messages: Internal standard error
Rebooted DMS: same result.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 1200 / 7 May 2015

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740