

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 11/06/17 Instrument Serial Number: 100360
Instrument Location: AST/Cooper Landing, AST/Soldotna
Supervisor Name and Agency: Colleen O'Bryen, SCDL

Describe the Status Change or Issue with the Instrument:

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 11/06/17 @ 12:59 to AST/Soldotna
Out of Service Date/Time: 11/06/17 @ 10:32A from AST/CL

For Use by SCDL

Additional Notes

Instrument removed from AST/Cooper Landing
due to past closure + relocated to AST/
Soldotna as a replacement instrument.
COB 12/13/17

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Foster, Charles R (DPS)

From: Simmons, Ronny A (DPS)
Sent: Monday, November 06, 2017 1:31 PM
To: O'Bryant, Colleen S (DPS)
Cc: Wertanen, Matthew J (DPS); Foster, Charles R (DPS); Hershberger, Casey W (DPS); Zweifel, Michael A (DPS); Hughes, Maurice I (DPS); Gilmore, Dane R (DPS)
Subject: Soldotna Datamaster
Attachments: SKMBT_60117110613190.pdf

Follow Up Flag: Follow up
Flag Status: Flagged

Hello All, after confirming with Chuck that this good to go we have moved the Cooper Landing Datamaster to Soldotna AST. We are closing down the Cooper Landing Post. I attempted to do a VOC on the Soldotna DM SN:100350 which immediately gave me a detector overflow message and stopped.

A VOC was done on the Cooper Landing DM prior to pulling it. I have done a VOC and a test on myself with the Cooper Landing one after we got it to Soldotna and everything appeared to be working fine. I have emailed everyone that it is good to use again.

The Cooper Landing DM is serial number 100360.

Sgt. Hershberger will be brining stuff to the lab Wednesday already so I will have him bring up 100350. Thanks Ronny