

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100695

Received From: Anchorage PD

Date: 12/23/25

Analyst: Carlie Glaister

Reason for instrument return:

Internal standard errors on Subject tests.
Instrument also recently had a couple
blank errors on VOCs.

From the field

no case on 12/23/25
Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐