

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 05/15/2020
Effective: 05/15/2020

Version: DMEF 2020 R0
Status: Active

Instrument # 100365 *DN 06/14/21*
Start Date: 06/14/2021

Analysts: Derech walton

Reason for instrument evaluation

Internal standard errors

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Gas Valve upright or secured to wall | <input checked="" type="checkbox"/> All tubing and connections are secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 O-rings in replacement bag | |

Troubleshooting/Repair

Instrument has already been refurbished.

- VOC (attached)
- Internal Standard value $\leq 3.00\%$ *ND - 5.57%*
- Verify software version

- | | |
|---|---|
| <input type="checkbox"/> Return to Factory for Repair | <input checked="" type="checkbox"/> Ready for Calibration |
| <input type="checkbox"/> Ready for Certification | <input type="checkbox"/> Proceed to Close Out |

End Date: 06/14/21

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 06/14/2021

External Standard Test Values

Diagnostic Check

VERSIONS
DMT: 3.02 ✓
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C PASSED
(44.0°C - 52.0°C)
Breath Tube = 45.8°C PASSED
(38.0°C - 50.0°C)

SETTINGS

Lamp Voltage = 1.83 V PASSED
(1.13 V - 2.65 V)
Cooler Voltage = 1.84 V PASSED
(1.10 V - 2.58 V)
Bias Voltage = 80 V PASSED
(48 V - 112 V)
Chopper Freq = 512 Hz PASSED
(475 Hz - 575 Hz)
Barometer = 29.7 in

PUMP INFO

Flow Rate = 4.765 L/M PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO

PUMP ON PASSED
(0.001 V <= 0.010 V)
PUMP OFF PASSED
(0.002 V <= 0.010 V)

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

Xq = 0.103 (5.57%) (0.00% - 4.00%) FAILED

Diagnostic Check Failed