

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100367

Date: 01/24/2020
Time: 12:27:46

OPERATOR NAME:
INTOX I METERS
CERTIFICATE #: 1

Ca = 0.2000
CAL = 0.990191 0.800 <= CAL < 1.200
b1 = 0.0000 0.0000 <= b1 < 0.0040
b2 = 0.0049 0.0010 <= b2 < 0.0100
b3 = 0.0003 0.0000 <= b3 < 0.0040
Xq = 0.0816 0.0500 <= Xq < 0.2000
a21 = 1.171546 1.050 <= a21 < 1.300
a31 = 0.414106 0.300 <= a31 < 0.800