

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100676

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/31/2025

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.49: 0.079  
LOT #: 09523080A1  
EXPIRATION: 05/05/2025  
TANK PRESSURE: 1208 psi

BLANK TEST                      0.000    22:21  
INTERNAL STANDARD            VERIFIED   22:21

Gas Flow Error

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.5°C            PASSED  
Breath Tube      = 48.1°C            PASSED

### PUMP INFO

Flow Rate = 3.992 L/M               PASSED

### DETECTOR INFO

PUMP ON                               PASSED  
PUMP OFF                               PASSED

### FILTER INFO

Filter 1                                PASSED  
Filter 2                                PASSED  
Filter 3                                PASSED

### INTERNAL STANDARD

PASSED

File only  
CW 5/12/25