

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100368
Start Date: 11/6/23
Date Received: 10/23/23

Analyst(s): Carlie Glaister

Reason for instrument evaluation

Preventative maintenance

Visual Inspection

Clean

Cermetex Box (modem) is secure

Double Check Valve secured to wall

O-ring inspection and ensure 2 replacement O-rings in bag

Replace missing screws

Ensure modem cable present

All tubing and connections secure

As Found Conditions:

Barometric Pressure (inHg): 29.13 Temperature (C): 22.7 Relative Humidity (%): 24

Linearity Test (attached)

0.020g/210L Mnf: <u>Airgas</u>	Lot: <u>AG-217902</u>	Exp: <u>6/28/24</u>
0.080g/210L Mnf: <u>Airgas</u>	Lot: <u>AG-310901</u>	Exp: <u>4/19/25</u>
0.300g/210L Mnf: <u>Airgas</u>	Lot: <u>AG-217903</u>	Exp: <u>6/28/24</u>

Notes

Replaced O-ring. As found passed, recommending adjustment due to 0.080 & 0.300 standards reading close to 5%.

VOC (attached)

VOC Internal Standard value ≤ 3.00%

Verify software version

Non-Drinking Subject Test (attached)

Condition of item: Good Needs Repair

Ready for Adjustment

Ready for Calibration

End Date: 11/6/23

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

LINEARITY CHECK

Alaska Department of Public Safety
DATAMASTER dmt: 100368


Date: 11/06/2023
Time: 10:38:33

SUPERVISOR NAME:
CARLIE GLAISTER
OPERATOR'S #: 10436

TARGET AVERAGE STD. DEV.

0.020	0.0190	0.0000
0.078	0.0740	0.0000
0.293	0.2800	0.0007

SLOPE = 0.956614
 R^2 = 0.999997



OK 6

LINEARITY CHECK: Control Test 1

Alaska Department of Public Safety
DATAMASTER dmt: 100368


Date: 11/06/2023
Time: 10:38:35

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.020
TARGET AT 29.17: 0.020
LOT #: AG217902
EXPIRATION: 06/28/2024
TANK PRESSURE: 780 psi

BLANK TEST	0.000	10:39
INTERNAL STANDARD	VERIFIED	10:40
EXTERNAL STANDARD	0.019	10:40
BLANK TEST	0.000	10:41
EXTERNAL STANDARD	0.019	10:41
BLANK TEST	0.000	10:42
EXTERNAL STANDARD	0.019	10:42
BLANK TEST	0.000	10:43
EXTERNAL STANDARD	0.019	10:43
BLANK TEST	0.000	10:44
EXTERNAL STANDARD	0.019	10:44
BLANK TEST	0.000	10:45

Average = 0.0190
Std Dev = 0.0000



LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety
DATAMASTER dmt: 100368






Date: 11/06/2023
Time: 10:46:41

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.17: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 536 psi

BLANK TEST	0.000	10:47
INTERNAL STANDARD	VERIFIED	10:48
EXTERNAL STANDARD	0.074	10:48
BLANK TEST	0.000	10:49
EXTERNAL STANDARD	0.074	10:49
BLANK TEST	0.000	10:50
EXTERNAL STANDARD	0.074	10:50
BLANK TEST	0.000	10:51
EXTERNAL STANDARD	0.074	10:51
BLANK TEST	0.000	10:52
EXTERNAL STANDARD	0.074	10:52
BLANK TEST	0.000	10:53

Average = 0.0740
Std Dev = 0.0000



LINEARITY CHECK: Control Test 3

Alaska Department of Public Safety
DATAMASTER dmt: 100368


Date: 11/06/2023
Time: 10:55:27

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.300
TARGET AT 29.17: 0.293
LOT #: AG217903
EXPIRATION: 06/28/2024
TANK PRESSURE: 349 psi

BLANK TEST	0.000	10:56
INTERNAL STANDARD	VERIFIED	10:56
EXTERNAL STANDARD	0.281	10:57
BLANK TEST	0.000	10:58
EXTERNAL STANDARD	0.280	10:58
BLANK TEST	0.000	10:59
EXTERNAL STANDARD	0.280	10:59
BLANK TEST	0.000	11:00
EXTERNAL STANDARD	0.279	11:00
BLANK TEST	0.000	11:01
EXTERNAL STANDARD	0.280	11:01
BLANK TEST	0.000	11:02

Average = 0.2800
Std Dev = 0.0007



VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100368

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/06/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.17: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 544 psi

BLANK TEST	0.000	10:11
INTERNAL STANDARD	VERIFIED	10:11
Xq = 0.075	(0.92%)	
EXTERNAL STANDARD	0.073	10:11
X[1] = 0.0734	(-0.0015)	(-0.0017)
BLANK TEST	0.000	10:12
EXTERNAL STANDARD	0.074	10:12
X[1] = 0.0738	(-0.0006)	(-0.0005)
BLANK TEST	0.000	10:13
EXTERNAL STANDARD	0.074	10:13
X[1] = 0.0738	(-0.0010)	(-0.0009)
BLANK TEST	0.000	10:14
EXTERNAL STANDARD	0.074	10:15
X[1] = 0.0742	(-0.0011)	(-0.0007)
BLANK TEST	0.000	10:15
EXTERNAL STANDARD	0.074	10:16
X[1] = 0.0738	(-0.0010)	(-0.0002)
BLANK TEST	0.000	10:17

Average = 0.0738
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C (44.0°C - 52.0°C)	PASSED
Breath Tube = 45.3°C (38.0°C - 50.0°C)	PASSED

SETTINGS

Lamp Voltage = 1.64 V (0.97 V - 2.27 V)	PASSED
Cooler Voltage = 1.71 V (1.04 V - 2.42 V)	PASSED
Bias Voltage = 79 V (47 V - 111 V)	PASSED
Chopper Freq = 526 Hz (475 Hz - 575 Hz)	PASSED
Barometer = 29.2 in	

PUMP INFO

Flow Rate = 4.590 L/M (3.500 L/M - 6.500 L/M)	PASSED
--	--------

DETECTOR INFO


PUMP ON (0.002 V <= 0.010 V)	PASSED
PUMP OFF (0.002 V <= 0.010 V)	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD Xq = 0.075 0.34% (0.00% - 4.00%)	PASSED
---	--------

40.00



SUBJECT TEST

Alaska Department of Public Safety
DATAMASTER dmt: 100368Date: 11/06/2023
Time: 11:26:01

SUBJECT NAME:

TEST TEST
LICENSE #: XXXX

OPERATOR NAME:

CARLIE GLAISTER
OPERATOR'S #: 10436DEPT/AGENCY:
NON-DRINKING SUBJECT TEST

CASE/REPORT: XXXX

TEST TYPE: TEST

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.17: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025

DIAGNOSTIC CHECK	PASSED	11:27
BLANK TEST	0.000	11:27
INTERNAL STANDARD	VERIFIED	11:27
EXTERNAL STANDARD	0.074	11:28
BLANK TEST	0.000	11:29
SUBJECT SAMPLE (Vol=1.61L)	0.000	11:29
BLANK TEST	0.000	11:30
EXTERNAL STANDARD	0.075	11:30
BLANK TEST	0.000	11:31

CVG