

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 5/13/2025

Instrument #: 100670

Instrument Location: AST/Mat-Su West

Supervisor Name and Agency: Colleen O'Bryant

Describe the Status Change or Issue with the Instrument:

Click or tap here to enter text.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: N/A

Out of Service Date/Time: N/A

For Use by SCDL

Additional Notes

While signing VOCs from 2025.04.11 VOC day Charles Foster discovered the dry gas external standard lot number for AST/Mat-Su West DataMaster DMT was incorrectly entered during a tank change on 2025.04.01. This is the same time the instrument was placed in service. The incorrect lot number is AG3335202, Exp 2025.12.18. The correct lot number is AG335202, Exp 2025.012.18. I contacted contacted AST/MSW on 2025.05.13 and confirmed the lot number of the tank and requested the agency perform a VOC to correct the stored lot number.

All subject tests run on DataMaster 100670 from 2025.04.01 to 2025.05.13 that list a dry gas external standard lot number of AG3335202, Exp 2025.12.18 were actually run utilizing the approved external standard lot number of AG335202, EXP 2025.12.18. This administrative error has no affect on the accuracy of the instrument.

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Email completed form to dps.scdl.toxicology@alaska.gov.