

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Acceptance Form – DMAF2011R0		
	Issuing Authority – Nita Boiz, Scientific Director		
	Approval date: 03/01/11	Effective date: 03/01/11	Archive date: Active

DataMaster DMT Acceptance Form

Instrument # 100422Date: 10/7/11Analyst: CRF

1. Check in instrument in to JusticeTrax Breath Alcohol Database ☐ N/A *Already Performed*
2. Visual Inspection ☒
3. Update software ☒
4. Print calibration factors (attached) Printed Previously
5. NPAS calibration paperwork (attached)
6. Barcode Instrument ☒
7. State of Alaska Property ID# N/A JBER

Completing Acceptance that
began on 8/19/11 by Colleen O'Bryant



	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Acceptance Form -- DMAF2011R0		
	Issuing Authority -- Nita Bolz, Scientific Director		
Approval date: 03/01/11		Effective date: 03/01/11	Archive date: Active

DataMaster DMT Acceptance Form

Instrument # 100422Date: 8/19/11Analyst: Colleen O'Bryant

1. Check in instrument in to JusticeTrax Breath Alcohol Database ☒
2. Visual Inspection ☒
3. Update software ☐
4. Print calibration factors (attached)
5. NPAS calibration paperwork (attached)
6. Barcode Instrument ☐
7. State of Alaska Property ID# _____

Check in not completed. Software not updated.
NPAS paperwork did not come with instrument
when delivered from JBE.

Returned to NPAS - interference with external
standard.

Replaced w/ new sticker + updated instrument
serial number
08/19/11



old SN

Page 1 of 1

Case: B-100422 By O'Bryant, Colleen - 08/19/2011

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100422

Date: 08/15/2011
Time: 18:31:02

SUPERVISOR NAME:
COLLEEN S O'BRYANT
CERTIFICATE #: 6237

LOT #: GUTH100
BOTTLE #: 11050
EXPIRATION: 04/06/2013

Ca = 0.1000
CAL = 0.991964 0.800 <= CAL < 1.200
b1 = 0.0001 0.0000 <= b1 < 0.0040
b2 = 0.0056 0.0010 <= b2 < 0.0100
b3 = 0.0002 0.0000 <= b3 < 0.0040
Xg = 0.0960 0.0500 <= Xg < 0.2000
a21 = 1.229637 1.050 <= a21 < 1.300
a31 = 0.414175 0.300 <= a31 < 0.800

Final page for report

\

\ANCSCLSAN01\JTRAX\CASEDOCS\B-100422\B-100422 ECF-20140

7/9/2014