

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100421

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/18/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.53: 0.079

LOT #: AG335202

EXPIRATION: 12/18/2025

TANK PRESSURE: 839 psi

BLANK TEST	0.000	15:30
INTERNAL STANDARD	VERIFIED	15:30
EXTERNAL STANDARD	0.079	15:30
BLANK TEST	0.000	15:31
EXTERNAL STANDARD	0.079	15:31

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C PASSED

Breath Tube = 47.2°C PASSED

PUMP INFO

Flow Rate = 4.729 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only jmw 12/19/25