

VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument
State of Alaska*

Serial #: 100368

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/10/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.54: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 1193 psi

BLANK TEST	0.000	21:25
INTERNAL STANDARD	VERIFIED	21:26
EXTERNAL STANDARD	INTERFERENCE	21:26

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber =	49.0°C	PASSED
Breath Tube =	40.2°C	PASSED

PUMP INFO

Flow Rate =	4.515 L/M	PASSED
-------------	-----------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED