

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100666

Date: 10/27/15

Analyst: CTF

1. Reason for instrument evaluation

Return from NPAS Repair due to Message pop-ups

2. Visual Inspection

3. Troubleshooting/Repair

Updated software from 2.00 → 3.00 → 3.01

Issues described that led to NPAS repair are no longer occurring.

4. VOC (attached)

5. Internal Standard value $\leq 3.00\%$ Yes No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100666

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/27/2015

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.48: 0.079
LOT #: 01315080A2
EXPIRATION: 02/05/2017
TANK PRESSURE: 1040 psi

BLANK TEST	0.000	11:55
INTERNAL STANDARD	VERIFIED	11:55
Xq = 0.086 (1.31%)		
EXTERNAL STANDARD	0.079	11:56
X[1] = 0.0788 (-0.0038) (-0.0009)		
BLANK TEST	0.000	11:57
EXTERNAL STANDARD	0.080	11:57
X[1] = 0.0796 (-0.0032) (-0.0008)		
BLANK TEST	0.000	11:58
EXTERNAL STANDARD	0.079	11:58
X[1] = 0.0792 (-0.0032) (-0.0008)		
BLANK TEST	0.000	11:59
EXTERNAL STANDARD	0.079	11:59
X[1] = 0.0795 (-0.0028) (-0.0009)		
BLANK TEST	0.000	12:00
EXTERNAL STANDARD	0.080	12:00
X[1] = 0.0796 (-0.0032) (-0.0006)		
BLANK TEST	0.000	12:01

Average = 0.0794
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.01
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 42.7°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.59 V	PASSED
(0.97 V - 2.27 V)	
Cooler Voltage = 1.61 V	PASSED
(0.98 V - 2.28 V)	
Bias Voltage = 79 V	PASSED
(48 V - 112 V)	
Chopper Freq = 518 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.5 in	

PUMP INFO

Flow Rate = 4.775 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.003 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.085	PASSED
2.31% (0.00% - 4.00%)	

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 20 _____