

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100695

Received From: Sand Point PD

Date: 5/19/25 - 5/21/2025

Analyst: Carlie Glaister

### Reason for instrument return:

Lack of agency communication w/ crime lab.  
1 file uploaded - 4/11/25 VOC. Records indicate  
missing datafiles. See 100695-2025-05-21-Change-In-Status  
Instrument returned w/ tank inside.

### From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

### From Intoximeters

Intoximeters Repair Paperwork (attached) ☐

Change instrument and case location in JusticeTrax Breath Alcohol Database ☐

Reset Options and Clear Memory ☐