

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100684

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/26/2017

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.23: 0.075
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 434 psi

BLANK TEST	0.000	15:24
INTERNAL STANDARD	VERIFIED	15:25
EXTERNAL STANDARD	INTERFERENCE	15:25

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 47.3°C	PASSED

PUMP INFO

Flow Rate = 3.826 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED