

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100369

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 02/14/2025

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

|                         |        |
|-------------------------|--------|
| Sample Chamber = 49.0°C | PASSED |
| Breath Tube = 46.8°C    | PASSED |

|                       |        |
|-----------------------|--------|
| PUMP INFO             |        |
| Flow Rate = 4.622 L/M | PASSED |

|               |        |
|---------------|--------|
| DETECTOR INFO |        |
| PUMP ON       | PASSED |
| PUMP OFF      | PASSED |

RFI DETECTED

File only  
02/16/25