

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 10/30/2021 Instrument Serial Number: 100360

Instrument Location: Sand Point Police Department

Supervisor Name and Agency: Chief David E. Anderson, SPPD

Describe the Status Change or Issue with the Instrument:

At approximately 0750 hours on the aforementioned date after entering subjects information into the Datamaster the machine began to cycle and then displayed, Internal standard error, followed by fatal system error. The machine then shut down to black screen. I attempted to run a second test and the machine shut down again. I returned several hours later and was able to run two normal tests.

The unit was not placed out of service.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740