

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:09/05/2025

Serial #: 100345

External Standard Test Values	Diagnostic Check
Filter Wheel Error	
	VERSIONS
	DMT: 3.02
	PIC: 3.02
	Modem: 2.6
	Questions: 2.2
	TEMPERATURES
	Sample Chamber = 48.7°C
	Breath Tube = 48.1°C
	PASSED
	PASSED
	PUMP INFO
	Flow Rate = 4.102 L/M
	PASSED
	DETECTOR INFO
	PUMP ON
	PUMP OFF
	PASSED
	PASSED
	FILTER INFO
	Filter 1
	Filter 2
	Filter 3
	PASSED
	PASSED
	PASSED
	INTERNAL STANDARD
	PASSED

File Only JMW 12/19/25