

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100415

Received From: APD

Date: 1/8/25

Analyst: Megan Carlock

Reason for instrument return:

Fatal errors and ticket db error

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒ Removed SD card, retrieved VOC from 12/27/24, uploaded, then replaced the SD card, upload worked as normal

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐