

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100676

Received From: Yakutat PD

Date: 12/9/24

Analyst: Megan Carlock

Reason for instrument return:

Damaged breath hose and hose holder. The plastic white piece where a mouthpiece is fitted is unattached from the rest of the hose.

Change instrument and case location in JusticeTrax Breath Alcohol Database ☒

Ensure change in instrument status form is in JusticeTrax ☒

Upload ☒

Update agency information and polling status in DMHost Calling Program ☒

Check for VOCs ☒

From Intoximeters

Intoximeters Repair Paperwork (attached) ☐