

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 12/31/2024

Instrument #: 100352

Instrument Location: North Slope Borough PD

Supervisor Name and Agency: Carlie Glaister - SCDL

Describe the Status Change or Issue with the Instrument:

This instrument has not been able to connect/poll since 8/16/2024. I spoke to breath test supervisor Christopher Dominey over the phone to try to troubleshoot the instrument connection. Christopher explained to me that he has been working on this issue and that the modem line was changed by the telecommunications company at some point to a fax/internet line (VOIP). This type of line is not compatible with the DataMaster. He is aware of this and is attempting to get it changed back to a fax line without internet/VOIP. However, this has not been able to be accomplished at this time. The communication issue does not impact the DataMaster's ability to perform VOCs and breath tests, and the instrument is in service at this time.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

-----  
**For Use by SCDL**

**Additional Notes**

Click or tap here to enter text.

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).**

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor