

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 11/5/2018 Instrument Serial Number: 100384

Instrument Location: Hoonah PD

Supervisor Name and Agency: Michael Ivie Hoonah DPS

Describe the Status Change or Issue with the Instrument:

Data Master is back online and new external standard has been put in. VOC was conducted and Data Master is good to go.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 11/1/2018 18:30

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

Instrument was originally attempted to be placed in service on 9/8/18 there were series of failed VOCs that were followed by a few passing. The tank was changed on 10/31/18 and the instrument placed in service as listed above. No tests were run prior to 11/1/18. See attached email.

BMB 11/15/18

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

Foster, Charles R (DPS)

From: Foster, Charles R (DPS)
Sent: Tuesday, October 30, 2018 10:15 AM
To: 'mike ivie'; 'jhuskey@cityofhoonah.org'
Cc: O'Bryant, Colleen S (DPS); Walton, Derek J (DPS); Barnett, Brandi M (DPS)
Subject: RE: New gas tank

We're sending you a new tank this week. Given the high number of failed VOCs starting in September, I am not going to sign the few that passed. Take your DMT out of service until you install the new tank. Our database shows that you have not used the DMT for any casework during that timeframe.

Thank you,
Chuck

From: mike ivie [mailto:rescueswimmer_99@yahoo.com]
Sent: Friday, October 26, 2018 2:59 AM
To: Foster, Charles R (DPS)
Subject: New gas tank

Hey Chuck, this is Officer Ivie with Hoonah PD. I keep getting a Standard out of Range message when I do VOC. I was wondering if you could send us a new tank to see if that would fix the problem. Our tank now is a 585 psi and it expires on 12/5/2018.

Thank you,
Michael Ivie