

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100342

Received From Intoximeters

Date 02/03/21

Analyst Derek Watten

1. Reason for instrument return:

Refurbished by manufacturer for gas flow errors

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

*Instrument was not in the field jmw 02/03/21*

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)

**SRO Number:** SRO-043146

**Received Date:** 11/24/2020

**Completion Date:** 01/07/2021

**SRO Type:** REPAIR

**SRO Description:** DMT Repair

**Customer ID:** C000003651

**Customer PO:** WTY03

**Contact:** Derek Walton

**Email:** derek.walton@alaska.gov

**Ship To:**

Alaska Department of Public Safety  
Procurement & Supply Section  
4805 Dr. MLK Jr. Avenue  
Anchorage AK 99507

**Units on SRO**

1 LEGDMT LEGACY DMT

Service Line: 1		<b>**All Instruments are Calibrated to Factory Specifications**</b>	
<b>Unit Type:</b>	(LEGDMT) LEGACY DMT		
<b>Serial Number:</b>	50100342		
<b>Reason:</b>	DMT Status Messages		
	GasFloEr		
<b>Reason Notes:</b>	Customer Reported Issue: experiencing persistent gas flow errors		
<b>Resolution:</b>	Replaced		
	RPL mech		
<b>Resolution Notes:</b>	Replaced the Dry Gas regulator Recalibrated touch screen Adjusted tech screen settings Adjusted RFI sensitivity Instrument Calibrated to Factory Specifications Accuracy tests passed Leaked checked - 24 hours  Parts: 001217-00 Dry Gas regulator  Calibration solution (0.200) Lot#20160 Bottle#1147 Dry gas: AG909102, Tank 032, Exp. 01 Apr 2021		
<b>Parts:</b>	<u>Qty</u>	<u>Part Number</u>	<u>Part Description</u>
	1	001217-00	DMT HRDW GAS MOUNT ASSY W/VALVE

# Order Packing Slip Report

**Intoximeters**  
Experience • Service • Integrity

1/18/2021 12:45:35 PM

**Packing Slip:** 219235

**Pack Date:** 1/18/2021

**SRO:** SRO-043146



**From:** Intoximeters, Inc  
2081 Craig Road  
St Louis MO 63146  
USA

**Customer:** C000003651

**Ship To:** 2

Alaska Department of Public Safety  
4805 Dr MLK Jr Ave  
Anchorage AK 99507  
USA

Alaska Department of Public Safety  
Procurement & Supply Section  
4805 Dr. MLK Jr. Avenue  
Anchorage AK 99507  
USA

**Customer PO:** WTY03 112520-  
Barnett

**Ship Via:** FDX

FED EX

**Weight:** 38.00

**Delivery Terms :** 2D

**FOB :** DESTINATION

**# Packages:** 1

**Customer Ship Account :**

Line Number	Item	UM	Ordered Quantity	Shipment Quantity	Backorder Quantity
1	LEGDMT	LEGACY DMT			50100342
1	001217-00	EA	1.00		0.00

DMT HRDW GAS MOUNT ASSY W/VALVE

*Rec'd 02/03/21  
JM*