VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument State of Alaska

Serial #: 100377 Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program Date: 03/13/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.13: 0.078 LOT #: AG735001

EXPIRATION: 12/16/2019 TANK PRESSURE: 89 psi

BLANK TEST

0.000 10:12

INTERNAL STANDARD VERIFIED 10:12

Tank Pressure Too Low

Diagnostic Check

VERSIONS DMT: 3.02

PIC: 3.03 Modem: 2.6 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C Breath Tube = 46.4°C PASSED PASSED

PUMP INFO Flow Rate = 4.675 L/M

PASSED

DETECTOR INFO

PASSED

PUMP ON PUMP OFF PASSED

FILTER INFO

Filter 1 Filter 2 Filter 3

PASSED PASSED PASSED

INTERNAL STANDARD

PASSED