

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 12 May 2025

Instrument #: 12- 22146 / 100365

Instrument Location: Eielson AFB, AK

Supervisor Name and Agency: SSgt. Malarie Zach - Eielson AFB JMW 06/26/25

Describe the Status Change or Issue with the Instrument:

Instrument would show interference detected even after clearing the area and running again.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA^①

Out of Service Date/Time: NA^①

For Use by SCDL

Additional Notes

① Instrument never placed in service at agency due to interference detected and internal standard error status messages JMW 06/26/25

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor