

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100670

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/15/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.08: 0.078
LOT #: AG528101
EXPIRATION: 10/08/2027
TANK PRESSURE: 1220 psi

BLANK TEST	0.000	09:13
INTERNAL STANDARD	VERIFIED	09:13
EXTERNAL STANDARD	0.078	09:14
BLANK TEST	0.000	09:15
EXTERNAL STANDARD	0.078	09:15
BLANK TEST	0.000	09:16
EXTERNAL STANDARD	0.078	09:16
BLANK TEST	0.000	09:17
EXTERNAL STANDARD	0.078	09:17
BLANK TEST	0.000	09:18
EXTERNAL STANDARD	INTERFERENCE	09:18

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 4.673 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only DM 12/22/25