

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100361

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/14/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.67: 0.079

LOT #: 09214080A3

EXPIRATION: ~~12/16/2019~~ 05/01/2016 Bomb 7/26/19

TANK PRESSURE: 633 psi

BLANK TEST	0.000	02:53
INTERNAL STANDARD	VERIFIED	02:53
EXTERNAL STANDARD	0.078	02:53
BLANK TEST	0.000	02:54
EXTERNAL STANDARD	0.079	02:55
BLANK TEST	0.000	02:56
EXTERNAL STANDARD	0.078	02:56
BLANK TEST	0.000	02:57
EXTERNAL STANDARD	0.078	02:57
BLANK TEST	0.000	02:58
EXTERNAL STANDARD	0.078	02:58
BLANK TEST	0.000	02:59

Average = 0.0782

Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C

PASSED

Breath Tube = 43.0°C

PASSED

PUMP INFO

Flow Rate = 4.073 L/M

PASSED

DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

FILTER INFO

Filter 1

PASSED

Filter 2

PASSED

Filter 3

PASSED

INTERNAL STANDARD

PASSED