

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100347

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/05/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.44: 0.073
LOT #: AG528101
EXPIRATION: 10/08/2027
TANK PRESSURE: 1221 psi

BLANK TEST	0.000	13:15
INTERNAL STANDARD	VERIFIED	13:15
EXTERNAL STANDARD	0.073	13:15
BLANK TEST	0.000	13:16
EXTERNAL STANDARD	0.073	13:16
BLANK TEST	0.000	13:17
EXTERNAL STANDARD	0.073	13:18
BLANK TEST	0.000	13:18
EXTERNAL STANDARD	0.074	13:19
BLANK TEST	0.000	13:20
EXTERNAL STANDARD	0.073	13:20
BLANK TEST	0.000	13:21

Average = 0.0732
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C
Breath Tube = 46.8°C

PASSED
PASSED

PUMP INFO
Flow Rate = 4.230 L/M

PASSED

DETECTOR INFO
PUMP ON
PUMP OFF

PASSED
PASSED

FILTER INFO
Filter 1
Filter 2
Filter 3

PASSED
PASSED
PASSED

INTERNAL STANDARD

PASSED


I, Charles R. Foster, after being first duly sworn, depose and state as follows:

- (1) I am a current employee of the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) As of the date signed below, I certify that the calibration of the referenced instrument was accurate on the date in which the above tests were performed and therefore certified for evidentiary use in the State of Alaska.



Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of 01, 20 20



Notary Public
My Commission Expires With Office



Tech Reviewer Initials: JW

Date: 12/19/25