

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100423  
Start Date: 5/17/23  
Date Received: 5/17/23

Analyst(s): C. O'Bryant

### Reason for instrument evaluation

Various status messages + reports of a subject  
spitting into the breath hose.

### Visual Inspection

- |   |   |
|---|---|
| <input checked="" type="checkbox"/> Clean (not including breath hose)                         | <input checked="" type="checkbox"/> Replace missing screws            |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure                            | <input checked="" type="checkbox"/> Ensure modem cable present        |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall                        | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag |   |

### As Found Conditions:

Barometric Pressure (inHg): \_\_\_\_\_ Temperature (C): COB Relative Humidity (%) \_\_\_\_\_

### Linearity Test (attached)

0.020g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_  
0.080g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_  
0.300g/210L Mnf: \_\_\_\_\_ Lot: \_\_\_\_\_ Exp: \_\_\_\_\_

### Notes

Breath hose has a visible blockage on the netting  
inside the white plastic opening where the  
mouthpiece is attached. Could not perform VOC.

- VOC (attached) COB  Verify software version  
 VOC Internal Standard value < 3.00% COB  Non-Drinking Subject Test (attached) COB

Condition of item:  Good  Needs Repair

- Ready for Adjustment  
 Ready for Calibration

End Date: 5/17/23

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100423

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/17/2023

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.95: 0.080  
LOT #: 09523080A1  
EXPIRATION: 05/05/2025  
TANK PRESSURE: 1023 psi

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 11:57 |
| INTERNAL STANDARD | VERIFIED | 11:57 |
| EXTERNAL STANDARD | 0.077    | 11:57 |
| BLANK TEST        | 0.000    | 11:58 |
| EXTERNAL STANDARD | 0.075    | 11:58 |
| BLANK TEST        | 0.000    | 11:59 |
| EXTERNAL STANDARD | 0.074    | 12:00 |

Standard Out of Range

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7 °C      PASSED  
Breath Tube = 48.1 °C      PASSED

PUMP INFO  
Flow Rate = 4.478 L/M      PASSED

DETECTOR INFO  
PUMP ON      PASSED  
PUMP OFF      PASSED

FILTER INFO  
Filter 1      PASSED  
Filter 2      PASSED  
Filter 3      PASSED

INTERNAL STANDARD      PASSED