

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100348

Analyst(s): Derek Walton

Start Date: 11/09/22

Date Received: 09/07/22

### Reason for instrument evaluation

Preventative maintenance due

### Visual Inspection

- |   |   |
|---|---|
| <input checked="" type="checkbox"/> Clean   | <input checked="" type="checkbox"/> Replace missing screws            |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure                            | <input checked="" type="checkbox"/> Ensure modem cable present        |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall                        | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag |   |

### As Found Conditions:

Barometric Pressure (inHg): 29.84 Temperature (C): 21.8 Relative Humidity (%): 21

### Linearity Test (attached)

0.020g/210L Mnf: <u>Airgas</u>	Lot: <u>A6103401</u>	Exp: <u>01/04/23</u>
0.080g/210L Mnf: <u>Airgas</u>	Lot: <u>A613402</u>	Exp: <u>05/14/23</u>
0.300g/210L Mnf: <u>Airgas</u>	Lot: <u>A6130502</u>	Exp: <u>11/01/23</u>

### Notes

Replaced o-ring, Tightened source block

- |   |  |
|---|--|
| <input checked="" type="checkbox"/> VOC (attached)                      | <input checked="" type="checkbox"/> Verify software version              |
| <input checked="" type="checkbox"/> VOC Internal Standard value ≤ 3.00% | <input checked="" type="checkbox"/> Non-Drinking Subject Test (attached) |

Condition of item:  Good  Needs Repair

- |   |
|---|
| <input type="checkbox"/> Ready for Adjustment             |
| <input checked="" type="checkbox"/> Ready for Calibration |

End Date: 11/09/22

### LINEARITY CHECK

Alaska Department of Public Safety  
DATAMASTER dmt: 100348


Date: 11/09/2022  
Time: 14:32:58

SUPERVISOR NAME:  
DEREK WALTON  
OPERATOR'S #: 9508

TARGET AVERAGE STD. DEV.

0.020	0.0196	0.0005
0.078	0.0760	0.0000
0.294	0.2860	0.0000

SLOPE = 0.972252  
 $R^2$  = 1.000000



DW

**LINEARITY CHECK: Control Test 1**

Alaska Department of Public Safety  
DATAMASTER dmt: 100348


Date: 11/09/2022  
Time: 14:33:01

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.020  
TARGET AT 29.35: 0.020  
LOT #: AG100401  
EXPIRATION: 01/04/2023  
TANK PRESSURE: 468 psi

BLANK TEST	0.000	14:34
INTERNAL STANDARD	VERIFIED	14:34
EXTERNAL STANDARD	0.019	14:34
BLANK TEST	0.000	14:35
EXTERNAL STANDARD	0.019	14:35
BLANK TEST	0.000	14:36
EXTERNAL STANDARD	0.020	14:36
BLANK TEST	0.000	14:37
EXTERNAL STANDARD	0.020	14:37
BLANK TEST	0.000	14:38
EXTERNAL STANDARD	0.020	14:38
BLANK TEST	0.000	14:39

Average = 0.0196  
Std Dev = 0.0005



# LINEARITY CHECK: Control Test 2

Alaska Department of Public Safety  
DATAMASTER dmt: 100348


Date: 11/09/2022  
Time: 14:41:17

## EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.35: 0.078  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 920 psi

BLANK TEST	0.000	14:42
INTERNAL STANDARD	VERIFIED	14:42
EXTERNAL STANDARD	0.076	14:42
BLANK TEST	0.000	14:43
EXTERNAL STANDARD	0.076	14:43
BLANK TEST	0.000	14:44
EXTERNAL STANDARD	0.076	14:45
BLANK TEST	0.000	14:45
EXTERNAL STANDARD	0.076	14:46
BLANK TEST	0.000	14:47
EXTERNAL STANDARD	0.076	14:47
BLANK TEST	0.000	14:48

Average = 0.0760  
Std Dev = 0.0000



**LINEARITY CHECK: Control Test 3**

Alaska Department of Public Safety  
DATAMASTER dmt: 100348


Date: 11/09/2022  
Time: 14:55:17

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.300  
TARGET AT 29.35: 0.294  
LOT #: AG130502  
EXPIRATION: 11/01/2023  
TANK PRESSURE: 423 psi

BLANK TEST	0.000	14:56
INTERNAL STANDARD	VERIFIED	14:56
EXTERNAL STANDARD	0.286	14:56
BLANK TEST	0.000	14:57
EXTERNAL STANDARD	0.286	14:58
BLANK TEST	0.000	14:58
EXTERNAL STANDARD	0.286	14:59
BLANK TEST	0.000	15:00
EXTERNAL STANDARD	0.286	15:00
BLANK TEST	0.000	15:01
EXTERNAL STANDARD	0.286	15:01
BLANK TEST	0.000	15:02

Average = 0.2860  
Std Dev = 0.0000



# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100348

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 11/09/2022

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.36: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023  
TANK PRESSURE: 896 psi

BLANK TEST	0.000	15:07
INTERNAL STANDARD ✓	VERIFIED	15:08
Xq = 0.091	(2.25%)	
EXTERNAL STANDARD	0.077	15:08
X[1] = 0.0775 (-0.0008) (0.0026)		
BLANK TEST	0.000	15:09
EXTERNAL STANDARD	0.077	15:09
X[1] = 0.0768 (-0.0005) (0.0025)		
BLANK TEST	0.000	15:10
EXTERNAL STANDARD	0.076	15:10
X[1] = 0.0762 (-0.0014) (0.0019)		
BLANK TEST	0.000	15:11
EXTERNAL STANDARD	0.076	15:11
X[1] = 0.0764 (-0.0006) (0.0019)		
BLANK TEST	0.000	15:12
EXTERNAL STANDARD	0.077	15:12
X[1] = 0.0765 (-0.0011) (0.0026)		
BLANK TEST	0.000	15:13

Average = 0.0766  
Std Dev = 0.0005

## Diagnostic Check

### VERSIONS

DMT: 3.02 ✓  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7 °C (44.0 °C - 52.0 °C)	PASSED
Breath Tube = 43.8 °C (38.0 °C - 50.0 °C)	PASSED

### SETTINGS

Lamp Voltage = 1.71 V (1.01 V - 2.35 V)	PASSED
Cooler Voltage = 1.71 V (1.04 V - 2.44 V)	PASSED
Bias Voltage = 79 V (48 V - 112 V)	PASSED
Chopper Freq = 537 Hz (475 Hz - 575 Hz)	PASSED
Barometer = 29.4 in	

### PUMP INFO

Flow Rate = 4.421 L/M (3.500 L/M - 6.500 L/M)	PASSED
--	--------

### DETECTOR INFO

PUMP ON (0.002 V <= 0.010 V)	PASSED
PUMP OFF (0.001 V <= 0.010 V)	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD ✓ Xq = 0.091 1.76% (0.00% - 4.00%)	PASSED
---	--------

**SUBJECT TEST**

Alaska Department of Public Safety  
DATAMASTER dmt: 100348

Date: 11/09/2022  
Time: 15:18:55

SUBJECT NAME:  
TEST TEST  
LICENSE #: XXXX

OPERATOR NAME:  
DEREK WALTON  
OPERATOR'S #: 9508

DEPT/AGENCY:  
NON-DRINKING SUBJECT  
CASE/REPORT: XXXX  
TEST TYPE: TEST

**EXTERNAL STANDARD INFORMATION**

NOMINAL: 0.080  
TARGET AT 29.36: 0.079  
LOT #: AG113402  
EXPIRATION: 05/14/2023

DIAGNOSTIC CHECK	PASSED	15:19
BLANK TEST	0.000	15:20
INTERNAL STANDARD	VERIFIED	15:20
EXTERNAL STANDARD	0.076	15:20
BLANK TEST	0.000	15:21
SUBJECT SAMPLE (Vol=2.92L)	0.000	15:21
BLANK TEST	0.000	15:22
EXTERNAL STANDARD	0.077	15:22
BLANK TEST	0.000	15:23

DM

