

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 6/24/15 Instrument Serial Number: 100391

Instrument Location: Wrangell PD

Supervisor Name and Agency: Bruce Smith

Describe the Status Change or Issue with the Instrument:

Bruce emailed regarding an interference detected
on an external standard.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: _____

For Use by SCDL

Additional Notes

I looked into this instrument and this is the first
occurrence of interference detected on an external
standard for the life of this tank, which was
installed in August 2014. I told Bruce to let me
know if this becomes a recurring issue.
See attached email. CRB 6/24/15

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O`Bryant, Colleen S (DPS)

From: O`Bryant, Colleen S (DPS)
Sent: Wednesday, June 24, 2015 9:19 AM
To: Bruce Smith (wpdsgt@aptalaska.net)
Subject: RE: Wrangell Instrument 100391

Bruce,

It looks like this is the first instance of an interference detected for that specific external standard, which was installed in August of 2014.

Thank you for letting us know. Please let me know if this becomes a recurring issue.

Colleen O`Bryant – Forensic Scientist III

From: DPS SCDL Toxicology (DPS sponsored)
Sent: Monday, June 22, 2015 8:40 AM
To: O`Bryant, Colleen S (DPS); Foster, Charles R (DPS); Barnett, Brandi M (DPS); Bolz, Nita J (DPS)
Subject: FW: Wrangell Instrument 100391

From: Bruce Smith [wpdsgt@aptalaska.net]
Sent: Friday, June 19, 2015 6:06 PM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: Wrangell Instrument 100391

I am letting you know that Wrangell's DataMaster 100391 just had a first time Interference status message from a second External Standard result on a non-drinking subject test. The following test was without issue.

Sergeant Bruce Smith
Wrangell Police Department