

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100342

Received From: Unalaska PD

Date: 3/23/26

Analyst: Kaja Carlson

Reason for instrument return:

Missing data files. 12/12/25 VOC needed.
20 files recovered. TicketDb issue resolved with
removal of SD card.

From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

Change instrument and case location in JusticeTrax Breath Alcohol Database

Reset Options and Clear Memory