

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument  
State of Alaska

Serial #: 100385

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 4/14/2017

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.91: 0.080  
LOT #: AG608504  
EXPIRATION: 3/25/2018  
TANK PRESSURE: 1118 psi

BLANK TEST	0.000	23:52
INTERNAL STANDARD	VERIFIED	23:52
EXTERNAL STANDARD	0.080	23:52
BLANK TEST	0.000	23:53
EXTERNAL STANDARD	0.080	23:53
BLANK TEST	0.000	23:54
EXTERNAL STANDARD	0.081	23:54
BLANK TEST	0.000	23:55
EXTERNAL STANDARD	0.080	23:55
BLANK TEST	0.000	23:56
EXTERNAL STANDARD	0.080	23:57
BLANK TEST	0.000	23:57

Average = 0.0802  
Std Dev = 0.0004

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED  
Breath Tube = 42.9°C PASSED

### PUMP INFO

Flow Rate = 4.411 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

### FILTER INFO


Filter 1 PASSED  
Filter 2 PASSED  
Filter 3 PASSED

### INTERNAL STANDARD

PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

  
Charles R. Foster  
Scientific Director  
State Breath Alcohol Program

5/8/17

Subscribed and sworn before me this 8<sup>th</sup> day of May, 20 17

  
Nikki Roth, Notary Public  
My Commission Expires With Office



CAB 4/26/17