

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 3/20/2025

Version: 4.0

Instrument #: 100358

Received From: AST TOK

Date: 3/23/26

Analyst: Kariya Carlson

Reason for instrument return:

Internal standard errors. Not uploading via Modem
line at lab. Uploaded via SD card.

From the field

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)

Change instrument and case location in JusticeTrax Breath Alcohol Database

Reset Options and Clear Memory