

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 11/10/2021

Version: 2.0

Instrument #: 100384

Analyst(s): C.O.Bryant

Start Date: 8/17/22

Reason for instrument evaluation

Returned to SCOD to reconcile files.

Visual Inspection

- | | |
|---|---|
| <input checked="" type="checkbox"/> Clean | <input checked="" type="checkbox"/> Replace missing screws |
| <input checked="" type="checkbox"/> Cermetex Box (modem) is secure | <input checked="" type="checkbox"/> Ensure modem cable present |
| <input checked="" type="checkbox"/> Double Check Valve secured to wall | <input checked="" type="checkbox"/> All tubing and connections secure |
| <input checked="" type="checkbox"/> O-ring inspection and ensure 2 replacement O-rings in bag | |

Troubleshooting/Repair

Replaced a cracked o-ring on the regulator.

- | | |
|--|--|
| <input checked="" type="checkbox"/> VOC (attached) | <input checked="" type="checkbox"/> Verify software version |
| <input checked="" type="checkbox"/> VOC Internal Standard value \leq 3.00% | <input checked="" type="checkbox"/> Non-Drinking Subject Test (attached) |
| <input type="checkbox"/> Return to Factory | <input type="checkbox"/> Ready for Calibration |
| <input checked="" type="checkbox"/> Ready for Certification | <input type="checkbox"/> Proceed to Close Out |
| <input type="checkbox"/> Further Evaluation Needed | |

End Date: 8/17/22

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100384

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/17/2022

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.18: 0.078
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 839 psi

BLANK TEST	0.000	15:04
INTERNAL STANDARD	VERIFIED	15:04
Xq = 0.090	(0.48%)	
EXTERNAL STANDARD	0.076	15:04
X[1] = 0.0761	(-0.0019)	(0.0023)
BLANK TEST	0.000	15:05
EXTERNAL STANDARD	0.076	15:06
X[1] = 0.0762	(-0.0013)	(0.0023)
BLANK TEST	0.000	15:07
EXTERNAL STANDARD	0.077	15:07
X[1] = 0.0770	(-0.0006)	(0.0019)
BLANK TEST	0.000	15:08
EXTERNAL STANDARD	0.077	15:08
X[1] = 0.0765	(-0.0013)	(0.0019)
BLANK TEST	0.000	15:09
EXTERNAL STANDARD	0.077	15:09
X[1] = 0.0769	(-0.0003)	(0.0022)
BLANK TEST	0.000	15:10

Average = 0.0766
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber	= 48.9 °C	PASSED
(44.0 °C - 52.0 °C)		
Breath Tube	= 48.1 °C	PASSED
(38.0 °C - 50.0 °C)		

SETTINGS

Lamp Voltage	= 1.62 V	PASSED
(0.98 V - 2.30 V)		
Cooler Voltage	= 1.83 V	PASSED
(1.07 V - 2.49 V)		
Bias Voltage	= 80 V	PASSED
(48 V - 112 V)		
Chopper Freq	= 526 Hz	PASSED
(475 Hz - 575 Hz)		
Barometer	= 29.2 in	

PUMP INFO

Flow Rate	= 4.698 L/M	PASSED
(3.500 L/M - 6.500 L/M)		

DETECTOR INFO


PUMP ON		PASSED
(0.002 V <= 0.010 V)		
PUMP OFF		PASSED
(0.002 V <= 0.010 V)		

FILTER INFO

Filter 1		PASSED
Filter 2		PASSED
Filter 3		PASSED

INTERNAL STANDARD

Xq = 0.091 0.39% (0.00% - 4.00%) PASSED



SUBJECT TEST	
Alaska Department of Public Safety	
DATAMASTER dmt: 100384	
TEST TEST	
LICENSE #: NON-DRINKING	
SUBJECT NAME:	
TEST TEST	
OPERATOR NAME:	
COLLEEN S O'BRYANT	
OPERATOR'S #: 6237	
DEPT/AGENCY:	
AAA1	
CASE/REPORT: 22-1234	
TEST TYPE: TEST	
EXTERNAL STANDARD INFORMATION	
NOMINAL: 0.080	
TARGET AT 29.18: 0.078	
LOT #: AGI13402	
EXPIRATION: 05/14/2023	
DIAGNOSTIC CHECK	PASSED
BLANK TEST	0.000
INTERNAL STANDARD	VERIFIED
EXTERNAL STANDARD	
BLANK TEST	0.077
SUBJECT SAMPLE (Vol=2.58L)	0.000
BLANK TEST	0.000
EXTERNAL STANDARD	0.077
BLANK TEST	0.000

100384 2022.08.17 Evaluation

Page 3 of 3