

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100695

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/14/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.59: 0.079

LOT #: AG436502

EXPIRATION: 12/30/2026

TANK PRESSURE: 977 psi

BLANK TEST	0.000	02:29
INTERNAL STANDARD	VERIFIED	02:29
EXTERNAL STANDARD	0.079	02:29
BLANK TEST	0.000	02:30
EXTERNAL STANDARD	0.079	02:30
BLANK TEST	0.000	02:31
EXTERNAL STANDARD	INTERFERENCE	02:31

Blank Error

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED

Breath Tube = 44.3°C PASSED

PUMP INFO

Flow Rate = 4.308 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only jmw 12/22/25