Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Effective: 10/10/2022	Version: 3.0
Instrument #: <u>/00385</u> Analys Start Date: <u>/0/14/シ</u> Date Received: <u>/0/14/シ</u>	st(s): <u>C. OBryan t</u>
Reason for instrument evaluation Unable to load ticket ISn't responding.	db, touch Ocreen
Visual Inspection X Clean	X Replace missing screws
X Cermetex Box (modem) is secure	X Ensure modem cable present
Nouble Check Valve secured to wall	old X All tubing and connections secure
X O-ring inspection and ensure 2 replacer	ment O-rings in bag
Linearity Test (attached) 0.020g/210L Mnf: 0.080g/210L Mnf: 0.300g/210L Mnf: Lot:	Exp: Exp: Exp:
Notes Valle to perform au fou Hicket als mrg.	
USC (attached) USC Internal Standard value ≤ 3.00% □	Verify software version Non-Drinking Subject Test (attached)
Condition of item: Good Needs Rep	pair
Ready for AdjustmentReady for Calibration	
End Date: 10/14/22	
Page 1 o All printed copies are uncontrolled.	of 1 Approved by: Chemistry Superviso
All printed copies are uncontrolled.	Approved by a chemistry Supervis

.