Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form Issued: 1/13/2017 Version: DMCIF 2017 R0 Effective: 1/17/2017 Status: Active Instrument S/N 100669 Received From Interimeters Analyst Deren Watton Date 10/23/21 1. Reason for instrument return: Refurbished by factory for Filter which emors 2. Change instrument and case location in JusticeTrax Breath Alcohol Database 3. Ensure change in instrument status form is in JusticeTrax Instrument was not in the field DM 10/23/20 4. Upload 5. Update agency information and polling status in DM Host Calling Program 6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)



Experience · Service · Integrity

Service Record

SRO Number: SRO-042069 Received Date: 09/16/2020 Completion Date: 09/28/2020

SRO Type: REPAIR SRO Description: DMT Repair/Preventive Maint

Customer ID: C000003651 Customer PO: 09162020-Walton

Contact: Derek Walton Email: derek.walton@alaska.gov

Ship To: Units on SRO

Alaska Department of Public Safety Procurement & Supply Section 4805 Dr. MLK Jr. Avenue Anchorage AK 99507

and replaced discolored tubing.

001199-00 59-0036-00 1 LEGDMT LEGACY DMT

Service Line:	**All Instruments are Calibrated to Factory Specifications**						
Unit Type:	(LEGDMT)	LEGACY DMT					
Serial Number:	: 50100669						
Reason:	DMT Status Messages						
	FltWl er						
Reason Notes:	Customer Reported Issue: Filter wheel errors						
Resolution:	Prev M						
	PrevMain						
	Removed filter wheel assembly for cleaning and lubrication. Cleaned & adjusted filter wheel solenoid. Replaced filter wheel solenoid mounting screws. Replaced stepper motor adjustment lock down screw. Replaced filter wheel motor pinion. Cleaned, lubricated, & adjusted filter wheel moving parts.						
Parts:	Qty	Part Number	Part Description				
	1	90-8000-00	CONTRACT PREV MAINT DMT				
	Preventive Maintenance						
Reason:	Preventive l	Maintenance					
Reason:	Preventive I	Maintenance					
	planned	Maintenance eventive Maintenance (90-800)	0-00)				
	planned Contract Pro		0-00)				

Resolution Notes: No leaks detected from the dry gas regulator. Sealed the sample chamber mirror blocks with metal foil tape. Inspected 5-Way valve

DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN

DMT HRDW CLIP BREATH TUBE HOLD 37058

Order Packing Slip Report

Intoximeters
Experience · Service · Integrity

Pack Date: 10/15/2020

10/15/2020 9:11:47 AM

Packing Slip: 213318

SRO: SRO-042069

From: Intoximeters, Inc

2081 Craig Road St Louis MO 63146

USA

Customer: C000003651

Ship To: 2

Alaska Department of Public Safety

4805 Dr MLK Jr Ave Anchorage AK 99507

USA

Alaska Department of Public Safety Procurement & Supply Section 4805 Dr. MLK Jr. Avenue Anchorage AK 99507

USA

Customer PO: 09162020-Walton

Ship Via: FDX

FED EX

Weight: 37.00

Delivery Terms: 2D

FOB: ORIGIN

Packages: 1

Customer Ship Account:

Line Number	Item	UM	Ordered Quantity	Shipment Quantity:	Backorder Quantity				
1	LEGDMT	LEGACY	DMT		50100669				
1	90-8000-00	EA		1.00	0.00				
CONTRACT PREV MAINT DMT									
2	001199-00	EA		1.00	0.00				
DMT ELEC REGULATOR V LDO 3.3V OUT 30V IN									
3	59-0036-00	EA		1.00	0.00				
DMT HRDW CLIP BREATH TUBE HOLD 37058									

CALIBRATION FACTORS

Alaska Department of Public Safety

DATAMASTER dmt: 100669

Date: 09/25/2020 Time: 11:59:08

OPERATOR NAME:

ST LOUIS INTOX A TECH CERTIFICATE #: 0U812

Ca = 0.1000

CAL	=	1.004968	0.800	<=	CAL	<	1.200
b1	=	0.0003	0.0000	<=	b1	<	0.0040
b2	=	0.0028	0.0010	<=	b2	<	0.0100
b3	=	0.0003	0.0000	<=	b3	<	0.0040
Χq	=	0.0855	0.0500	<=	рX	<	0.2000
a21	=	1.169427	1.050	<=	a21	<	1.300
a31	=	0.518949	0.300	<=	a31	<	0.800